
Submission ID: P08.VZ6.1D2.64G

Part 6: Substance Specific Requirements, relating to safety-engineered devices for all hollow bore needles

1. Name: William Lorne Boudreau RN

(a) Feedback is provided: on own behalf

Organization:

(b) Feedback is provided: from a worker's perspective

2. E-mail:

3. Feedback:

I do not feel safe using the new plastic cover style of needle protection that swings into place. The device due to the cumbersome design and the fact that it can and does come open after the safety cover has been deployed is not the highest level of protection available. The retractable or spring loaded devices are much safer since once the device is activated you cannot accidentally re-expose the needle without disassembling the device.

4. Please indicate your level of support of the proposed amendments:
generally agree