
Submission ID: P08.Z86.HQ2.7L0

Part 6: Substance Specific Requirements, relating to safety-engineered devices for all hollow bore needles

1. Name: paul terpstra

(a) Feedback is provided: on own behalf

Organization:

(b) Feedback is provided: from a worker's perspective

2. E-mail:

3. Feedback:

section 6: We need safety engineered sharps in the workplace but not just the cheapest ones on the market. The retractable needle ones are the best. The sheathed ones are more dangerous.

4. Please indicate your level of support of the proposed amendments:
generally agree